

K06 3/89

Toshiba America Medical Systems, Inc.
510(k) Pre-market Notification

510(k) Summary

Date: September 28, 2006 NOV - 3 2006

Submitter's Name: Toshiba America Medical Systems, Inc.

Submitter's Address: P.O. Box 2068, 2441 Michelle Drive,
Tustin, CA 92781-2068

Submitter's Contact: Paul Biggins, Regulatory Affairs Specialist,
(714)730-5000

**Establishment Registration
Number:** 2020563

Device Proprietary Name: TSX-301A, Aquilion ONE CT System

Common Name: Scanner, Computed Tomography, X-Ray
[Fed. Reg. No. 892.1750, Pro. Code: 90JAK]

Regulatory Class: II (per 21 CFR 892.1750)

Performance Standard: 21 CFR Subchapter J,
Federal Diagnostic X-ray Equipment Standard

Predicate Device(s): TSX-101A/I Aquilion 64 SP CT System

Reason For Submission New device

Description of this Device:

The TSX-301A/1 is a whole body CT scanner. This device produces axial images that may be reconstructed into 3-D volumes. The device consists of a gantry, patient couch (table) and peripheral cabinets used for data processing and display.

Summary of Intended Uses:

This device is intended to produce cross sectional data sets of the anatomy.

Technological Characteristics:

This device employs similar materials and processes as found in the predicate device. The device produces ionizing radiation that is employed to generate cross sectional images of the anatomy.

Toshiba America Medical Systems, Inc.
510(k) Pre-market Notification

Safety and Effectiveness Concerns:

This device is designed and manufactured under the Quality System Regulations as outlined in 21 CFR § 820. All requirements of the Federal Diagnostic Equipment Standard, as outlined in 21 CFR § 1020.30 and 1020.33, that apply to this upgrade, will be met and reported via a supplement to the initial report for the predicate device. Additionally this system is in conformance with the applicable parts of the IEC 60601-1 {applicable portions}; IEC 60601-2-32, and IEC 60601-2-44. - Medical Device Safety standards .

Substantial Equivalence:

This device is similar in materials and processes to that of the predicate device, Toshiba TSX-101A/H Aquilion 64 SP CT scanner.



DEPARTMENT OF HEALTH & HUMAN SERVICES

Food and Drug Administration
9200 Corporate Blvd.
Rockville MD 20850

Toshiba America Medical Systems, Inc.
c/o Mr. Mark Job
Responsible Third Party Official
Regulatory Technology Services LLC
1394 25th Street, N.W.
BUFFALO MN 55313

NOV - 3 2006

Re: K063189
Trade/Device Name: TSX-301A, Aquilion ONE CT System
Regulation Number: 21 CFR §892.1750
Regulation Name: Computed tomography x-ray system
Regulatory Class: II
Product Code: JAK
Dated: October 16, 2006
Received: October 20, 2006

Dear Mr. Job:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (Premarket Approval), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.



Protecting and Promoting Public Health

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at one of the following numbers, based on the regulation number at the top of this letter:

21 CFR 876.xxx	(Gastroenterology/Renal/Urology	240-276-0115
21 CFR 884.xxx	(Obstetrics/Gynecology)	240-276-0115
21 CFR 894.xxx	(Radiology)	240-276-0120
Other		240-276-0100

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (240) 276-3150 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,



Nancy C. Brogdon
Director, Division of Reproductive,
Abdominal, and Radiological Devices
Office of Device Evaluation
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known): K063189

Device Name: TSX-301A, Aquilion ONE CT System

Indications For Use:

Acquisition and display of axial x-ray images of the whole body to include the head.

Prescription Use ☒
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use ☐
(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Page 1 of _____

Monica Brogdon
(Division Sign-Off)
Division of Reproductive, Abdominal,
and Radiological Devices
510(k) Number K063189